

# Certificate of Analysis

**Product name** AMPLIRUN® ADENOVIRUS DNA CONTROL  
**Catalogue Number** MBC001  
**Lot Number** 24MBC001001-C  
**Expiry date** 2027-03-09

Kit contents	Lot Number	Expiry date
VIRCELL ADN DNA CONTROL	24MBC001101-C01	2027-03-09

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	17625 ± 1157
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

**Identity test oligonucleotides**

*Adeno-hexon-AQ1* GCCACGGTGGGGTTTCTAAACTT  
*Adeno-hexon-AQ2* GCCCCAGTGGTCTTACATGCACATC

Lieberman, D., et al. (2009). J Clin Microbiol, 47(11):3439-3443

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-C	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.  
 and fulfils the essential requirements as established in Directive 98/79/CE and part of Regulation (EU) 2017/746.  
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	16/09/2024	