

Certificate of Analysis

Product name	AMPLIRUN® SALMONELLA ENTERITIDIS DNA CONTROL
Catalogue Number	MBC003-R
Lot Number	24MBC003002-R
Expiry date	2026-09-26

Kit contents	Lot Number	Expiry date
VIRCELL SEN DNA CONTROL	24MBC003102-R01	2026-09-26

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	18000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides <i>S.enterica</i> F GGTAAGGGGCTTCGGTATC <i>S.enterica</i> R TATTGGCTCCCTGAATACGC <i>Lee, S. H., et al. (2009). J Vet Sci, 10(1):43-51</i>		
--	--	--

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-R	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	02/04/2024	