

Certificate of Analysis

Product name	AMPLIRUN® BACILLUS CEREUS DNA CONTROL
Catalogue Number	MBC004-R
Lot Number	23MBC004001-R
Expiry date	2025-11-02

Kit contents	Lot Number	Expiry date
VIRCELL BCE DNA CONTROL	23MBC004101-R01	2025-11-02

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	13000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

BCFomp2 CGCCTCGTTGGATGACG
BCRomp2 GATATACATTCACCTTGACTAATACCG

Oliwa-Stasiak, K., et al. (2011). Appl Environ Microbiol, 77(1):80-88

VIRCELL CONTROL RECONSTITUTION SOLUTION	22MBCRS002-R	2027-11-16
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Reca John, Isabel María	25/05/2023	

