

Certificate of Analysis

Product name AMPLIRUN® BARTONELLA HENSELAE DNA CONTROL
Catalogue Number MBC005-R
Lot Number 23MBC005001-R
Expiry date 2025-09-21

Kit contents	Lot Number	Expiry date
VIRCELL BHE DNA CONTROL	23MBC005101-R01	2025-09-21

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	16000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides
*Bartonella*ITS.F CTTCTGTTCTCTTTCTTCA
*Bartonella*ITS.R AACCAACTGAGCTACAAGCC
 Jensen, W. A., et al. (2000). J Clin Microbiol, 38(5):1717-1722

VIRCELL CONTROL RECONSTITUTION SOLUTION	22MBCRS002-R	2027-11-16
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Guerrero Ruiz, Juan José	29/03/2023	

