

Certificate of Analysis

Product name	AMPLIRUN® BARTONELLA QUINTANA DNA CONTROL
Catalogue Number	MBC006-R
Lot Number	25MBC006001-R
Expiry date	2027-09-17

Kit contents	Lot Number	Expiry date
VIRCELL BQU DNA CONTROL	25MBC006101-R01	2027-09-17

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	12041 ± 824
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

*Bartonella*ITS.F CTTCGTTTCTCTTTCTTCA

*Bartonella*ITS.R AACCAACTGAGCTACAAGCC

Jensen, W. A., et al. (2000). *J Clin Microbiol*, 38(5):1717-1722

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	21/03/2025	