

# Certificate of Analysis

Product name	AMPLIRUN® BORDETELLA PARAPERTUSSIS DNA CONTROL
Catalogue Number	MBC007-R
Lot Number	24MBC007001-R
Expiry date	2026-07-10

Kit contents	Lot Number	Expiry date
VIRCELL BPA DNA CONTROL	24MBC007101-R01	2026-07-10

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	16000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

**Identity test oligonucleotides**

Parap IS1001Fwd CCGGCTCGACGAATTGC  
 Parap IS1001Rev AGAACCAGAGCCGTTTGAGTT

Grogan, J.A., et al. (2011). J Med Microbiol, 60:722-729

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-R	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Reca John, Isabel María	17/01/2024	