

Certificate of Analysis

Product name AMPLIRUN® BRUCELLA ABORTUS DNA CONTROL
Catalogue Number MBC009-R
Lot Number 23MBC009001-R
Expiry date 2025-09-30


Kit contents	Lot Number	Expiry date
VIRCELL BAB DNA CONTROL	23MBC009101-R01	2025-09-30

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	14000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides
 BRU-UP GGGCAAGGTGGAAGATT
 BRU-LOW CGGCAAGGGTCGGTGT
Serpe, L., et al. (1999). J Dairy Res, 66(2):313-317

VIRCELL CONTROL RECONSTITUTION SOLUTION	22MBCRS002-R	2027-11-16
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Guerrero Ruiz, Juan José	13/04/2023	



Vircell, S.L.
 Parque Tecnológico de la Salud, Avicena 8, 18016 Granada, Spain T. +34 958 441 264
 info@vircell.com www.vircell.com