

# Certificate of Analysis

Product name	AMPLIRUN® CHLAMYDIA TRACHOMATIS DNA CONTROL
Catalogue Number	MBC012
Lot Number	24MBC012002-C
Expiry date	2026-12-03

Kit contents	Lot Number	Expiry date
VIRCELL CTR DNA CONTROL	24MBC012102-C01	2026-12-03

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	13338 ± 236
SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION
<b>Identity test oligonucleotides</b> Ctr_F CATGAAACTCGTTCCGAAATAGAA Ctr_R TCAGAGCTTACCTAACAACGCATA  <i>Jaton, K., et al. (2006). J Med Microbiol., 55:1667-1674</i>		

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-C	2028-10-13
---	--------------	------------

*This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. and fulfils the essential requirements as established in Directive 98/79/CE.*

*Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.*

## Q.C. Testing by

Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	11/06/2024	