

Certificate of Analysis

Product name	AMPLIRUN® CRYPTOCOCCUS NEOFORMANS DNA CONTROL
Catalogue Number	MBC015-R
Lot Number	23MBC015002-R
Expiry date	2025-12-12

Kit contents	Lot Number	Expiry date
VIRCELL CNE DNA CONTROL	23MBC015102-R01	2025-12-12

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	10000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides STR1F GAGATTCGGCAGGAAGAAGC STR1R CGTAAGGGATGACGAAAAGGTA Feng, X., et al. (2013). J Clin Microbiol, 51(6):1920-1923		
---	--	--

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-R	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Reca John, Isabel María	07/12/2023	