

Certificate of Analysis

Product name AMPLIRUN® COXIELLA BURNETII DNA CONTROL
Catalogue Number MBC018-R
Lot Number 23MBC018001-R
Expiry date 2025-12-13

Kit contents	Lot Number	Expiry date
VIRCELL CBU DNA CONTROL	23MBC018101-R01	2025-12-13


Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	18000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides
 COBU-icdpri_f: GACCGACCCATTATCCCT
 COBU-icdpri_r: CGGCGTAGATCTCCATCCA

De Bruin, A., et al. (2011). Appl Environ Microbiol., 77(18):6516-6523

VIRCELL CONTROL RECONSTITUTION SOLUTION	22MBCRS002-R	2027-11-16
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Reca John, Isabel María	21/06/2023	



Vircell, S.L.
 Parque Tecnológico de la Salud, Avicena 8, 18016 Granada, Spain T. +34 958 441 264
 info@vircell.com www.vircell.com