

Certificate of Analysis

Product name AMPLIRUN® COXIELLA BURNETII DNA CONTROL
Catalogue Number MBC018-R
Lot Number 24MBC018002-R
Expiry date 2027-05-21

Kit contents	Lot Number	Expiry date
VIRCELL CBU DNA CONTROL	24MBC018102-R01	2027-05-21

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	10428 ± 352
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

COBU-icdpri_f: GACCGACCCATTATCCCT

COBU-icdpri_r: CGGCGTAGATCTCCATCCA

De Bruin, A., et al. (2011). Appl Environ Microbiol., 77(18):6516-6523

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	02/12/2024	