

# Certificate of Analysis

Product name	AMPLIRUN® COXIELLA BURNETII DNA CONTROL
Catalogue Number	MBC018-R
Lot Number	24MBC018001-R
Expiry date	2025-12-13

Kit contents	Lot Number	Expiry date
VIRCELL CBU DNA CONTROL	24MBC018101-R01	2025-12-13

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	18000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

**Identity test oligonucleotides**

COBU-icdpri\_f: GACCGACCCATTATTCCT

COBU-icdpri\_r: CGGCGTAGATCTCCATCCA

De Bruin, A., et al. (2011). Appl Environ Microbiol., 77(18):6516-6523

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.  
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Martínez Ramírez, David	22/08/2024	