

Certificate of Analysis

Product name	AMPLIRUN® ENTEROVIRUS 71 RNA CONTROL
Catalogue Number	MBC019
Lot Number	23MBC019001-C
Expiry date	2025-06-19

Kit contents	Lot Number	Expiry date
VIRCELL ENT RNA CONTROL	23MBC019101-C01	2025-06-19

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	13000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

EV-EQ-F ACATGGTGCGAAGAGTCTATTGAGCT

EV-EQ-R CCAAAGTAGTCGGTTCCGC

Wong, K., et al. (2010). Appl Environ Microbiol, 76(19):6441-6448

VIRCELL CONTROL RECONSTITUTION SOLUTION	22MBCRS001-C	2027-11-16
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. and fulfils the essential requirements as established in Directive 98/79/CE.

Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Reca John, Isabel María	30/06/2023	

