

# Certificate of Analysis

Product name	AMPLIRUN® ENTEROVIRUS 71 RNA CONTROL
Catalogue Number	MBC019
Lot Number	25MBC019001-C
Expiry date	2027-01-13

Kit contents	Lot Number	Expiry date
VIRCELL ENT RNA CONTROL	25MBC019101-C01	2027-01-13

  

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	13466 ± 161
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

  

<b>Identity test oligonucleotides</b> EV-EQ-F ACATGGTGCGAAGAGTCTATTGAGCT EV-EQ-R CCAAAGTAGTCGGTTCCGC  Wong, K., et al. (2010). Appl Environ Microbiol, 76(19):6441-6448		
---	--	--

  

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-C	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. and fulfils the applicable general safety and performance requirements as established in Regulation (EU) 2017/746. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	03/02/2025	