

Certificate of Analysis

Product name	AMPLIRUN® ESCHERICHIA COLI (VTEC) DNA CONTROL
Catalogue Number	MBC022-R
Lot Number	25MBC022001-R
Expiry date	2027-11-20

Kit contents	Lot Number	Expiry date
VIRCELL ECO DNA CONTROL	25MBC022101-R01	2027-11-20

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	12125 ± 700
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

VTEC-VT1-FW CATTGTCTGGTGACAGTAGCT

VTEC-VT2-RV CCCGTAATTGCGCACTGAG

Bischoff, C., et al. (2005). J Microbiol Methods, 61(3): 335-341

VIRCELL CONTROL RECONSTITUTION SOLUTION	25MBCRS001-R	2030-01-20
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Tejera García, Hugo	29/05/2025	