

Certificate of Analysis

Product name AMPLIRUN® HERPES SIMPLEX 1 DNA CONTROL
Catalogue Number MBC023
Lot Number 23MBC023001-C
Expiry date 2025-07-19

Kit contents	Lot Number	Expiry date
VIRCELL HSV1 DNA CONTROL	23MBC023101-C01	2025-07-19


Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	19000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides
 HSV1-F GCAGTTTACGTACAACCACATACAGC
 HSV1&2-R AGCTTGCGGGCCTCGTT
 Namvar, L., et al. (2005). J Clin Microbiol, 43(5):2058-2064

VIRCELL CONTROL RECONSTITUTION SOLUTION	22MBCRS001-C	2027-11-16
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. and fulfils the essential requirements as established in Directive 98/79/CE.

Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Guerrero Ruiz, Juan José	24/01/2023	



Vircell, S.L.
Parque Tecnológico de la Salud, Avicena 8, 18016 Granada, Spain T. +34 958 441 264
info@vircell.com www.vircell.com