

Certificate of Analysis

Product name AMPLIRUN® HERPES SIMPLEX 2 DNA CONTROL
Catalogue Number MBC024
Lot Number 23MBC024001-C
Expiry date 2025-10-26

| Kit contents | Lot Number | Expiry date |
|--------------------------|-----------------|-------------|
| VIRCELL HSV2 DNA CONTROL | 23MBC024101-C01 | 2025-10-26 |


| Analysis | Specifications | Results |
|---------------------------------|---|---|
| QUANTIFICATION TEST (copies/μL) | 10000 - 20000 | 18000 |
| IDENTITY TEST | SPECIFIC AMPLIFICATION OF 1:1000 DILUTION | SPECIFIC AMPLIFICATION OF 1:1000 DILUTION |

Identity test oligonucleotides
 HSV2-F TGCAGTTTACGTATAACCACATACAGC
 HSV1&2-R AGCTTGCGGGCCTCGTT
 Namvar, L., et al. (2005). J Clin Microbiol, 43(5):2058-2064

| | | |
|---|--------------|------------|
| VIRCELL CONTROL RECONSTITUTION SOLUTION | 22MBCRS001-C | 2027-11-16 |
|---|--------------|------------|

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. and fulfils the essential requirements as established in Directive 98/79/CE.

Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

| Q.C. Testing by | | | |
|----------------------------|------------------------------|------------|---|
| Position | Name | Date | Signature |
| Quality Control Department | Gallardo Espigares, Cristina | 17/05/2023 |  |



Vircell, S.L.
 Parque Tecnológico de la Salud, Avicena 8, 18016 Granada, Spain T. +34 958 441 264
 info@vircell.com www.vircell.com