

Certificate of Analysis

Product name	AMPLIRUN® HERPES SIMPLEX 2 DNA CONTROL
Catalogue Number	MBC024
Lot Number	24MBC024001-C
Expiry date	2025-10-26

Kit contents	Lot Number	Expiry date
VIRCELL HSV2 DNA CONTROL	24MBC024101-C01	2025-10-26

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	18000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides HSV2-F TGCAGTTTACGTATAACCACATACAGC HSV1&2-R AGCTTGCGGGCCTCGTT <i>Namvar, L., et al. (2005). J Clin Microbiol, 43(5):2058-2064</i>		
--	--	--

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-C	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. and fulfils the essential requirements as established in Directive 98/79/CE.

Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	14/05/2024	