

Certificate of Analysis

Product name AMPLIRUN® HHV-6 DNA CONTROL
Catalogue Number MBC025-R
Lot Number 23MBC025002-R
Expiry date 2025-12-06


| Kit contents | Lot Number | Expiry date |
|--------------------------|-----------------|-------------|
| VIRCELL HHV6 DNA CONTROL | 23MBC025102-R01 | 2025-12-06 |

| Analysis | Specifications | Results |
|---------------------------------|---|---|
| QUANTIFICATION TEST (copies/μL) | 10000 - 20000 | 12000 |
| IDENTITY TEST | SPECIFIC AMPLIFICATION OF 1:1000 DILUTION | SPECIFIC AMPLIFICATION OF 1:1000 DILUTION |

Identity test oligonucleotides
 HHV6U41-F4: CGGAACATTGTTGAGCAGAAA
 HHV6U41-R106: AAGAAGAATCCCTTGTCTGGC
 Yavarian, J., et al. (2015). BMC Psychiatry, 15:296

| | | |
|---|--------------|------------|
| VIRCELL CONTROL RECONSTITUTION SOLUTION | 22MBCRS002-R | 2027-11-16 |
|---|--------------|------------|

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

| Q.C. Testing by | | | |
|----------------------------|-------------------------|------------|---|
| Position | Name | Date | Signature |
| Quality Control Department | Reca John, Isabel María | 15/06/2023 |  |



Vircell, S.L.
 Parque Tecnológico de la Salud, Avicena 8, 18016 Granada, Spain T. +34 958 441 264
 info@vircell.com www.vircell.com