

# Certificate of Analysis

**Product name** AMPLIRUN® HHV-6 DNA CONTROL  
**Catalogue Number** MBC025-R  
**Lot Number** 24MBC025001-R  
**Expiry date** 2025-12-06

Kit contents	Lot Number	Expiry date
VIRCELL HHV6 DNA CONTROL	24MBC025101-R01	2025-12-06

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	12000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

**Identity test oligonucleotides**

HHV6U41-F4: CGGAACATTGTTGAGCAGAAA

HHV6U41-R106: AAGAAGAATCCCTTGTCTGGC

*Yavarian, J., et al. (2015). BMC Psychiatry, 15:296*

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.  
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Martínez Ramírez, David	27/08/2024	