

Certificate of Analysis

Product name	AMPLIRUN® ROTAVIRUS RNA CONTROL
Catalogue Number	MBC026-R
Lot Number	24MBC026002-R
Expiry date	2026-08-22

Kit contents	Lot Number	Expiry date
VIRCELL ROT RNA CONTROL	24MBC026102-R01	2026-08-22

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	18567 ± 595
SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

RotaNSP3.F ACCATCTWCACRTRACCCTC

RotaNSP3.R GGTCACATAACGCCCCCTATA

Verheyen, J., et al. (2009). Appl Environ Microbiol, 75(9):2798-2801

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Martínez Ramírez, David	29/08/2024	