

# Certificate of Analysis

Product name	AMPLIRUN® YERSINIA ENTEROCOLITICA DNA CONTROL
Catalogue Number	MBC027-R
Lot Number	24MBC027001-R
Expiry date	2026-06-11

Kit contents	Lot Number	Expiry date
VIRCELL YEN DNA CONTROL	24MBC027101-R01	2026-06-11

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	18000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

**Identity test oligonucleotides**

foxA-F ACGGCGGTGATGTGAACAA

foxA-R GGGTCCACTTGCAGCACATT

Wang, J., et al. (2014). J Clin Microbiol, 52(12):4443-4444

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.  
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	05/09/2024	