

# Certificate of Analysis

**Product name** AMPLIRUN® INFLUENZA A H1 RNA CONTROL  
**Catalogue Number** MBC028  
**Lot Number** 24MBC028001-C  
**Expiry date** 2026-01-09

Kit contents	Lot Number	Expiry date
VIRCELL INAH1 RNA CONTROL	24MBC028101-C01	2026-01-09

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	14000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

**Identity test oligonucleotides**  
*Flu A Fw* GACCRATCCTGTACCTCTGAC  
*Flu A Rv* GGGCATTYTGACAAACGTCTACG  
*Chen, Y. et al. (2011). J Clin Microbiol, 49(4):1653-1656*

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-C	2028-10-13
---	--------------	------------

*This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. and fulfils the essential requirements as established in Directive 98/79/CE.*

*Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.*

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	19/01/2024	