

Certificate of Analysis

Product name AMPLIRUN® INFLUENZA A H3 RNA CONTROL
Catalogue Number MBC029
Lot Number 23MBC029001-C
Expiry date 2025-12-18

Kit contents	Lot Number	Expiry date
VIRCELL INAH3 RNA CONTROL	23MBC029101-C01	2025-12-18

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	20000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides
Flu A Fw GACCRATCCTGTACCTCTGAC
Flu A Rv GGGCATTYTGACAAKCGTCTACG
Chen, Y. et al. (2011). J Clin Microbiol, 49(4):1653-1656

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-C	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. and fulfils the essential requirements as established in Directive 98/79/CE.

Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	22/12/2023	