

Certificate of Analysis

Product name	AMPLIRUN® LEISHMANIA INFANTUM DNA CONTROL
Catalogue Number	MBC032-R
Lot Number	24MBC032001-R
Expiry date	2026-12-05

Kit contents	Lot Number	Expiry date
VIRCELL LIN DNA CONTROL	24MBC032101-R01	2026-12-05

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	10623 ± 84
SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

LCS-1 GCAATGCCAGCTACATATATG
LCS-3 CAGCTTTTGGGTGGGTAACA

da Silveira, Neto O.J., et al. (2012). Rev Bras Parasitol Vet, 21(3):304-307

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-R	2028-10-13
-----------------------------------------	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by

Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	12/06/2024	