

Certificate of Analysis

Product name	AMPLIRUN® MYCOBACTERIUM TUBERCULOSIS DNA CONTROL
Catalogue Number	MBC034
Lot Number	25MBC034001-C
Expiry date	2027-06-10

Kit contents	Lot Number	Expiry date
VIRCELL MTU DNA CONTROL	25MBC034101-C01	2027-06-10

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	15617 ± 1139
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

IS6110f: TCAAGGAGCACATCAGCC

IS6110r: TCACGGTTCAGGGTTAGC

Sanjuan-Jimenez, R., et al. (2013). PLoS ONE, 8(3):e58353

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-C	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. and fulfils the applicable general safety and performance requirements as established in Regulation (EU) 2017/746. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Tejera García, Hugo	20/01/2025	