

Certificate of Analysis

Product name	AMPLIRUN® MYCOPLASMA PNEUMONIAE DNA CONTROL
Catalogue Number	MBC035
Lot Number	24MBC035002-C
Expiry date	2027-04-29

Kit contents	Lot Number	Expiry date
VIRCELL MPN DNA CONTROL	24MBC035102-C01	2027-04-29

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	14808 ± 509
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides Mpn-3-identi.F TCTTCAGGCTCAGGTCAA Mpn-4-identi.R TTCCCGTATTAGTATTAGGC Pitcher, D., et al. (2006). J Med Microbiol, 55(Pt 2):149-55		
---	--	--

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-C	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. and fulfils the essential requirements as established in Directive 98/79/CE and part of Regulation (EU) 2017/746. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	07/11/2024	