

Certificate of Analysis

Product name	AMPLIRUN® PARAINFLUENZA 1 RNA CONTROL
Catalogue Number	MBC037
Lot Number	24MBC037001-C
Expiry date	2026-10-16

Kit contents	Lot Number	Expiry date
VIRCELL PAR1 RNA CONTROL	24MBC037101-C01	2026-10-16

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	15262 ± 342
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

PIS1 + CCGGTAATTCTCATACCTATG

PIS1 - CTTGGAGCGGAGTTGTTAAG

Echevarria, J.E., et al. (1998). J Clin Microbiol, 36(5):1388-1391

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-C	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. and fulfils the essential requirements as established in Directive 98/79/CE and part of Regulation (EU) 2017/746. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	25/10/2024	