

# Certificate of Analysis

**Product name** AMPLIRUN® PARAINFLUENZA 2 RNA CONTROL  
**Catalogue Number** MBC038  
**Lot Number** 25MBC038001-C  
**Expiry date** 2027-05-14

Kit contents	Lot Number	Expiry date
VIRCELL PAR2 RNA CONTROL	25MBC038101-C01	2027-05-14

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	18958 ± 959
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

**Identity test oligonucleotides**  
PIV2PR3 AACTATGTCCAGAGGAGAGGTGCTGG  
PIV2PR5 CCATGCCTGCATAAGCACACTGTAGC  
*Osiowy, C. (1998). J Clin Microbiol, 36(11):3149-3154*

VIRCELL CONTROL RECONSTITUTION SOLUTION	25MBCRS001-C	2030-05-08
---	--------------	------------

*This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. and fulfils the applicable general safety and performance requirements as established in Regulation (EU) 2017/746. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.*

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Tejera García, Hugo	22/05/2025	