

Certificate of Analysis

Product name AMPLIRUN® PARAINFLUENZA 3 RNA CONTROL
Catalogue Number MBC039
Lot Number 24MBC039001-C
Expiry date 2026-06-05

Kit contents	Lot Number	Expiry date
VIRCELL PAR3 RNA CONTROL	24MBC039101-C01	2026-06-05

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	15280 ± 749
SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

PIV3s GGAGCATTGTGTCATCTGTC

PIV3as TAGTGTGTAATGCAGCTCGT

Templeton, K. E., et al. (2004). J Clin Microbiol, 42(4):1564-1569

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-C	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
 and fulfils the essential requirements as established in Directive 98/79/CE.
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by

Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	13/06/2024	