

Certificate of Analysis

Product name AMPLIRUN® MUMPS RNA CONTROL
Catalogue Number MBC040-R
Lot Number 25MBC040001-R
Expiry date 2026-04-03

Kit contents	Lot Number	Expiry date
VIRCELL MUM RNA CONTROL	25MBC040101-R01	2026-04-03

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	17000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

MUMPS-SH61F GTGACCTGCCGTTGCA
 MUMPS-SH147R GTTATGATCAGAGAGAAGAATTAGCAATAG

Boddicker, J.D., et al. (2007). J Clin Microbiol, 45(9):2902-2908

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	27/01/2025	