

Certificate of Analysis

Product name	AMPLIRUN® RESPIRATORY SYNCYTIAL VIRUS(subtype A) RNA CONTROL
Catalogue Number	MBC041
Lot Number	25MBC041001-C
Expiry date	2027-05-05

Kit contents	Lot Number	Expiry date
VIRCELL RSVA RNA CONTROL	25MBC041101-C01	2027-05-05

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	19380 ± 125
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

RSVs GCCAAAAARTTTRTTCCACAATA

RSVas TCTTCATCWCCATACTTTCTGTGA

Syrmis, M.W., et al. (2004) J Mol Diagn., 6(2):125-31

VIRCELL CONTROL RECONSTITUTION SOLUTION	25MBCRS001-C	2030-05-08
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. and fulfils the applicable general safety and performance requirements as established in Regulation (EU) 2017/746. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Tejera García, Hugo	13/05/2025	