

# Certificate of Analysis

Product name	AMPLIRUN® RICKETTSIA CONORII DNA CONTROL
Catalogue Number	MBC042-R
Lot Number	23MBC042001-R
Expiry date	2026-06-12

Kit contents	Lot Number	Expiry date
VIRCELL RCO DNA CONTROL	23MBC042101-R01	2026-06-12

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	16000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

**Identity test oligonucleotides**

*Rick. PanR8-F* AGCTTGCTTTGGATCATTTGG  
*Rick. PanR8-R* TTCCTTGCCCTTTCATACATCTAGT

Balraj, P., et al (2009). Clin Microbiol Infect, 15(2):294-295

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-R	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	19/12/2023	