

Certificate of Analysis

Product name AMPLIRUN® RICKETTSIA CONORII DNA CONTROL
Catalogue Number MBC042-R
Lot Number 25MBC042001-R
Expiry date 2027-10-23

Kit contents	Lot Number	Expiry date
VIRCELL RCO DNA CONTROL	25MBC042101-R01	2027-10-23

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	16472 ± 543
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides
Rick.PanR8-F AGCTTGCTTTTGGATCATTTGG
Rick.PanR8-R TTCCTTGCCCTTTCATACATCTAGT
Balraj, P., et al (2009). Clin Microbiol Infect, 15(2):294-295

VIRCELL CONTROL RECONSTITUTION SOLUTION	25MBCRS001-R	2030-01-20
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	25/04/2025	