

Certificate of Analysis

Product name	AMPLIRUN® SALMONELLA TYPHI DNA CONTROL
Catalogue Number	MBC044-R
Lot Number	23MBC044002-R
Expiry date	2025-12-19

Kit contents	Lot Number	Expiry date
VIRCELL STY DNA CONTROL	23MBC044102-R01	2025-12-19

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	14000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides STY-0307F ATGAAACCTTTATTCTCAGTGC STY-0307R TTAGCGTAATTCCAGAACC Goay, Y. X., et al. (2016). Biomed Res Int. 8905675		
---	--	--

VIRCELL CONTROL RECONSTITUTION SOLUTION	22MBCRS002-R	2027-11-16
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	30/06/2023	

