

Certificate of Analysis

Product name	AMPLIRUN® MEASLES RNA CONTROL
Catalogue Number	MBC046-R
Lot Number	24MBC046001-R
Expiry date	2026-02-26

Kit contents	Lot Number	Expiry date
VIRCELL MEA RNA CONTROL	24MBC046101-R01	2026-02-26

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	14000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

MEAS-N1139-F TGGCATCTGAACTCGGTATCAC

MEAS-N1213-R TGTCTCAGTAGTATGCATTGCAA

Benamar, T., et al. (2016). PLoS ONE., 11 (1): e0147154

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-R	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	07/03/2024	