

Certificate of Analysis

Product name	AMPLIRUN® VARICELLA-ZOSTER VIRUS DNA CONTROL
Catalogue Number	MBC048
Lot Number	23MBC048002-C
Expiry date	2025-08-13

Kit contents	Lot Number	Expiry date
VIRCELL VZV DNA CONTROL	23MBC048102-C01	2025-08-13

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	11000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides


VZVORF38f CGGGTGAACCGTATTCTGAG
VZVORF38r TTGAACAATCACGAACCGTT

Tipples, G.A., et al. (2003). J Virol Methods, 113(2):113-116

VIRCELL CONTROL RECONSTITUTION SOLUTION	22MBCRS001-C	2027-11-16
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. and fulfils the essential requirements as established in Directive 98/79/CE.

Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Guerrero Ruiz, Juan José	22/02/2023	



Vircell, S.L.
Parque Tecnológico de la Salud, Avicena 8, 18016 Granada, Spain T. +34 958 441 264
info@vircell.com www.vircell.com