

Certificate of Analysis

| | |
|------------------|---|
| Product name | AMPLIRUN® PARAINFLUENZA 4 A RNA CONTROL |
| Catalogue Number | MBC050 |
| Lot Number | 24MBC050001-C |
| Expiry date | 2026-04-04 |

| Kit contents | Lot Number | Expiry date |
|---------------------------|-----------------|-------------|
| VIRCELL PAR4A RNA CONTROL | 24MBC050101-C01 | 2026-04-04 |

| Analysis | Specifications | Results |
|---------------------------------|---|---|
| QUANTIFICATION TEST (copies/μL) | 12500 - 20000 | 15000 |
| IDENTITY TEST | SPECIFIC AMPLIFICATION OF 1:1000 DILUTION | SPECIFIC AMPLIFICATION OF 1:1000 DILUTION |

Identity test oligonucleotides

HPIV-4 F AAAGAATTAGGTGCAACCAGTC


HPIV-4 R GTGTCTGATCCATAAGCAGC

Lau, S. K. P., et al. (2005). J Clin Microbiol, 43(9):4515-4521

| | | |
|---|--------------|------------|
| VIRCELL CONTROL RECONSTITUTION SOLUTION | 23MBCRS001-C | 2028-10-13 |
|---|--------------|------------|

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. and fulfils the essential requirements as established in Directive 98/79/CE.

Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

| Q.C. Testing by | | | |
|----------------------------|-------------------------|------------|---|
| Position | Name | Date | Signature |
| Quality Control Department | Conde Ortega, Ana María | 18/04/2024 |  |