

Certificate of Analysis

Product name	AMPLIRUN® NEISSERIA MENINGITIDIS SEROGROUP C DNA CONTROL
Catalogue Number	MBC053-R
Lot Number	23MBC053002-R
Expiry date	2026-04-25

Kit contents	Lot Number	Expiry date
VIRCELL NMEC DNA CONTROL	23MBC053102-R01	2026-04-25

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	14000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

NME ctrA F753 TGTGTTCCGCTATACGCCATT
NME ctrA R846 GCCATATTCACACGATATACC

Khater, W. S. & Elabd, S. H. (2016). Int J Microbiol, 4197187

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-R	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Reca John, Isabel María	30/10/2023	