

Certificate of Analysis

Product name	AMPLIRUN® ACANTHAMOEBA CASTELLANII DNA CONTROL
Catalogue Number	MBC054-R
Lot Number	23MBC054001-R
Expiry date	2025-12-26

Kit contents	Lot Number	Expiry date
VIRCELL ACA DNA CONTROL	23MBC054101-R01	2025-12-26

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	15000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides ACA-TaqAcF1 CGACCAGCGATTAGGAGACG ACA-TaqAcR1 CCGACGCCAAGGACGAC Chang, C.W. & Wu, Y.C. (2010). Water Sci Technol, 62(9):2141-2149		
---	--	--

VIRCELL CONTROL RECONSTITUTION SOLUTION	22MBCRS002-R	2027-11-16
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	25/07/2023	

