

Certificate of Analysis

Product name AMPLIRUN® DENGUE 1 VIRUS RNA CONTROL
Catalogue Number MBC055-R
Lot Number 24MBC055001-R
Expiry date 2026-06-25

Kit contents	Lot Number	Expiry date
VIRCELL DEN1 RNA CONTROL	24MBC055101-R01	2026-06-25

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	16327 ± 371
SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

DENV1_F CAATGGATGACAACAGAAGAYATG
 DENV1_R TCCATCCATGGGTTTCCTCTAT

Alm, E., et al. (2015). BMC Infect Dis., 15:493

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by

Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	28/06/2024	