

Certificate of Analysis

Product name	AMPLIRUN® DENGUE 1 VIRUS RNA CONTROL
Catalogue Number	MBC055-R
Lot Number	25MBC055001-R
Expiry date	2026-06-25

Kit contents	Lot Number	Expiry date
VIRCELL DEN1 RNA CONTROL	25MBC055101-R01	2026-06-25

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	16327 ± 371
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides DENV1_F CAATGGATGACAACAGAAGAYATG DENV1_R TCCATCCATGGGTTTCTCTAT Alm, E., et al. (2015).BMC Infect Dis.,15:493		
---	--	--

VIRCELL CONTROL RECONSTITUTION SOLUTION	25MBCRS001-R	2030-01-20
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Tejera García, Hugo	15/05/2025	