

# Certificate of Analysis

**Product name** AMPLIRUN® DENGUE 3 VIRUS RNA CONTROL  
**Catalogue Number** MBC057-R  
**Lot Number** 24MBC057001-R  
**Expiry date** 2026-07-24

Kit contents	Lot Number	Expiry date
VIRCELL DEN3 RNA CONTROL	24MBC057101-R01	2026-07-24

  

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	12665 ± 146
SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

  

<b>Identity test oligonucleotides</b> DEN-3 F GGACTGGACACACGCACTCA DEN-3 C CATGTCTCTACCTTCTCGACTTGTCT  <i>Johnson, B.W., et al. (2005). J Clin Microbiol, 43(10):4977-4983</i>		
--	--	--

  

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.  
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Martínez Ramírez, David	31/07/2024	