

Certificate of Analysis

Product name	AMPLIRUN® LEISHMANIA CHAGASI DNA CONTROL
Catalogue Number	MBC060-R
Lot Number	25MBC060001-R
Expiry date	2027-12-02

Kit contents	Lot Number	Expiry date
VIRCELL LCH DNA CONTROL	25MBC060101-R01	2027-12-02

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	13910 ± 254
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

LCS-1 GCAATGCCAGCTACATATATG

LCS-3 CAGCTTTTGGGTGGGTAACA

da Silveira, Neto O.J., et al. (2012). Rev Bras Parasitol Vet, 21(3):304-307

VIRCELL CONTROL RECONSTITUTION SOLUTION	25MBCRS001-R	2030-01-20
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Tejera García, Hugo	18/06/2025	