

# Certificate of Analysis

**Product name** AMPLIRUN® LEISHMANIA CHAGASI DNA CONTROL  
**Catalogue Number** MBC060-R  
**Lot Number** 23MBC060002-R  
**Expiry date** 2025-10-10

Kit contents	Lot Number	Expiry date
VIRCELL LCH DNA CONTROL	23MBC060102-R01	2025-10-10

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	14000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION


**Identity test oligonucleotides**

LCS-1 GCAATGCCAGCTACATATATG  
 LCS-3 CAGCTTTTGGGTGGGTAACA

*da Silveira, Neto O.J., et al. (2012). Rev Bras Parasitol Vet, 21(3):304-307*

VIRCELL CONTROL RECONSTITUTION SOLUTION	22MBCRS002-R	2027-11-16
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	05/09/2023	



Vircell, S.L.  
 Parque Tecnológico de la Salud, Avicena 8, 18016 Granada, Spain T. +34 958 441 264  
 info@vircell.com www.vircell.com