

Certificate of Analysis

Product name AMPLIRUN® COXSACKIE B5 RNA CONTROL
Catalogue Number MBC062-R
Lot Number 25MBC062001-R
Expiry date 2027-03-18

Kit contents	Lot Number	Expiry date
VIRCELL COB5 RNA CONTROL	25MBC062101-R01	2027-03-18

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	17817 ± 344
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

EV-EQ-F ACATGGTGCGAAGAGTCTATTGAGCT
 EV-EQ-R CCAAAGTAGTCGGTTCCGC

Wong, K., et al. (2010). Appl Environ Microbiol, 76(19):6441-6448

VIRCELL CONTROL RECONSTITUTION SOLUTION	25MBCRS001-R	2030-01-20
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	25/03/2025	