

# Certificate of Analysis

Product name	AMPLIRUN® ECHOVIRUS 5 RNA CONTROL
Catalogue Number	MBC064-R
Lot Number	23MBC064001-R
Expiry date	2025-10-10

Kit contents	Lot Number	Expiry date
VIRCELL ECH5 RNA CONTROL	23MBC064101-R01	2025-10-10

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	14000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

**Identity test oligonucleotides**

EV-EQ-F ACATGGTGCGAAGAGTCTATTGAGCT  
EV-EQ-R CCAAAGTAGTCGGTTCCGC

Wong, K., et al. (2010). Appl Environ Microbiol, 76(19):6441-6448

VIRCELL CONTROL RECONSTITUTION SOLUTION	23MBCRS001-R	2028-10-13
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Reca John, Isabel María	23/10/2023	

