

Certificate of Analysis

Product name	AMPLIRUN® NEISSERIA MENINGITIDIS SEROGROUP B DNA CONTROL
Catalogue Number	MBC066-R
Lot Number	23MBC066002-R
Expiry date	2025-12-27

Kit contents	Lot Number	Expiry date
VIRCELL NMEB DNA CONTROL	23MBC066102-R01	2025-12-27

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	18000
IDENTITY TEST	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION


Identity test oligonucleotides

NME ctrA F753 TGTGTTCCGCTATACGCCATT
NME ctrA R846 GCCATATTCACACGATATACC

Khater, W. S. & Elabd, S. H. (2016). Int J Microbiol, 4197187

VIRCELL CONTROL RECONSTITUTION SOLUTION	22MBCRS002-R	2027-11-16
-----------------------------------------	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L. Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by			
Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	26/07/2023	



Vircell, S.L.
Parque Tecnológico de la Salud, Avicena 8, 18016 Granada, Spain T. +34 958 441 264
info@vircell.com www.vircell.com