

Certificate of Analysis

Product name AMPLIRUN® WEST NILE VIRUS RNA CONTROL
Catalogue Number MBC069-R
Lot Number 24MBC069001-R
Expiry date 2026-07-18

Kit contents	Lot Number	Expiry date
VIRCELL WNV RNA CONTROL	24MBC069101-R01	2026-07-18

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	12500 - 20000	15433 ± 422
SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION

Identity test oligonucleotides

WNV E1 F CTACCGTCAGCGATCTC
WNV E1 R CTGGGTCAGCACGTTTG

Zink, S.D., et al. (2013). *Diagn Microbiol Infect Dis*, 77(2):129-132

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.
Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

Q.C. Testing by

Position	Name	Date	Signature
Quality Control Department	Conde Ortega, Ana María	26/07/2024	