

# Certificate of Analysis

Product name	AMPLIRUN® STREPTOCOCCUS PNEUMONIAE DNA CONTROL
Catalogue Number	MBC070-R
Lot Number	24MBC070001-R
Expiry date	2027-01-08

Kit contents	Lot Number	Expiry date
VIRCELL SPN DNA CONTROL	24MBC070101-R01	2027-01-08

Analysis	Specifications	Results
QUANTIFICATION TEST (copies/μL)	10000 - 20000	18143 ± 574
SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION	SPECIFIC AMPLIFICATION OF 1:1000 DILUTION
<b>Identity test oligonucleotides</b> SPN IytA F373    ACGCAATCTAGCAGATGAAGCA SPN IytA R424    TCGTGCGTTTAAATCCAGCT  <i>Wright, M. S., et al. (2017). Front Microbiol, 8:1661</i>		

VIRCELL CONTROL RECONSTITUTION SOLUTION	24MBCRS001-R	2029-06-03
---	--------------	------------

This is to certify that the above kit performs in accordance with the Quality Control specifications of Vircell, S.L.  
 Vircell, S.L. manufactures under ISO 9001 and ISO 13485 certified quality systems.

## Q.C. Testing by

Position	Name	Date	Signature
Quality Control Department	Martínez Ramírez, David	16/07/2024	